BALANCES & TEST SERVICE 2024





Analytical Balance KERN ABP-A - Features & Accessories of the Models



Automatic sliding doors, which can be opened and closed using sensors



Opening possible alternatively via button



Internal draft shield – minimises the effect of air currents in the weighing chamber, only for 0.01 mg models of the series ABP-A





KERN ABP-A with integrated ioniser and internal draft shield



loniser (included) can be activated at the touch of a button



Erlenmeyer piston holder, included as standard with ABP 200-5M and ABP 200-5AM



Multi-function weighing plate, Note: included with delivery for models with [d] = 0,01 mg



Model	Weighing	Readability	Verification	Minimal	Repro-	Linearity	Options	
	capacity		value	load	ducibility		Verification	DAkkS Calibr. Certificate
	[Max]	[d]	[e]	[Min]			M	DAkkS
KERN	g	mg	mg	mg	mg	mg	KERN	KERN
ABP 100-5AM	135	0,01	1	1	0,05	± 0,1	965-201	963-101
ABP 200-5AM	220	0,01	1	1	0,05	± 0,1	965-201	963-101
ABP 200-4AM	220	0,1	1	10	0,1	± 0,2	965-201	963-101
ABP 300-4AM	320	0,1	1	10	0,2	± 0,3	965-201	963-101
		Multi-d	ivision balance	e, with incre	easing or deci	reasing load, it sv	witches automatically	
			to the next lar	gest or sma	allest weighin	g range [Max] an	d readout [d].	
ABP 100-5DAM	52 120	0,01 0,1	1 1	1	0,02 0,1	± 0,05 0,2	965-201	963-101
ABP 200-5DAM	102 220	0,01 0,1	1 1	1	0,05 0,1	±0,1 0,2	965-201	963-101
Note: For devic	es that require	verification (c	onformity ass	essment ac	cording to NA	WI 2014/31/EU	J), please include the verificatio	n when placing your order.
T	The initial verific	cation is not p	ossible after d	elivery. Plea	ase inform the	e full address of t	the location of use for the initial	verification.





PREMIUM ★★★

BALANCES & TEST SERVICE 2024

Interface for second

second balance

Protocol (KCP)

It is a standardized

Network interface

an Ethernet network

KERN Communication


interface command set for

For direct connection of a


For connecting the scale to

balance


KERN Pictograms














Single cell technology Advanced version of the force compensation principle with the highest level of precision

Conformity Assessment Μ The time required for +3 DAYS conformity assessment is specified in the pictogram

DAkkS calibration DAkkS

possible (DKD) The time required for DAkkS calibration is shown in days in the pictogram



+3 DAYS

Factory calibration (ISO) The time required for Factory calibration is shown in days in the pictogram

Package shipment

The time required for internal shipping preparations is shown in days in the pictogram

Pallet shipment

The time required for 2 DAYS internal shipping preparations is shown in days in the pictogram

*The Bluetooth® word mark and logos are registered trademarks owned by Bluetooth SIG, Inc. and any use of such marks by KERN & SOHN GmbH is under license. Other trademarks and trade names are those of their respective owners



ANALOG

to connect a suitable

peripheral device for analogue processing of the measurements